

**AWG50CUP and AWG70CUP  
Upgrades for AWG5000C and AWG7000C  
Series Instruments**

**Instructions**

[www.tektronix.com](http://www.tektronix.com)



075-1009-00

**Tektronix**

Copyright © Tektronix. All rights reserved. Licensed software products are owned by Tektronix or its subsidiaries or suppliers, and are protected by national copyright laws and international treaty provisions.

Tektronix products are covered by U.S. and foreign patents, issued and pending. Information in this publication supersedes that in all previously published material. Specifications and price change privileges reserved.

TEKTRONIX and TEK are registered trademarks of Tektronix, Inc.

## **Contacting Tektronix**

Tektronix, Inc.  
14200 SW Karl Braun Drive  
P.O. Box 500  
Beaverton, OR 97077  
USA

For product information, sales, service, and technical support:

- In North America, call 1-800-833-9200.
- Worldwide, visit [www.tektronix.com](http://www.tektronix.com) to find contacts in your area.

---

# Kit Description

This kit provides instructions to install an option on an AWG5000C or AWG7000C series instrument. Once installed, the upgrade is equivalent to a factory-installed option.

For a list of available options, see the Tektronix Web site ([www.tektronix.com](http://www.tektronix.com)), or contact your Tektronix representative.

## Products

**AWG5000C and AWG7000C series B010100 and above**

## Kit Parts List

Option	Quantity	Part number	Description
All	1 each	NS <sup>1</sup>	LABEL, OPTION INFORMATION
All	1 each	075-0xxx-00	MANUAL, KIT INSTRUCTIONS

<sup>1</sup> NS – Not Salable

# Installation Instructions

To install your new option(s), you need the key number that is provided with this kit. This unique key number is valid only for the specific model number and serial number of your instrument; it will not work on any other instrument. This single key number enables all valid options for your instrument, those that were factory installed and any others that you may have already purchased. In other words, this new key number supersedes any previous key numbers that you may already have.



**CAUTION.** *To avoid losing your options, do not misplace the paper label included with this kit. This label contains the option installation key number and should be installed on the rear panel of the instrument for future reference.*

*As a backup, copy the key number and save it with the other important papers for your instrument. If you ever need to re-enable the option, you must reenter this key number.*

---

## Install the Option

Perform the following steps to enable your new option:

1. Power on the instrument.
2. From the menu bar, select the System menu, and then select **Option Installation**.
3. The Option Installation dialog box is displayed.




Figure 1: Option Installation dialog box (1)

4. Click **Continue** to display a second dialog box.



**Figure 2: Option Installation dialog box (2)**

5. To enable the new options, enter the alphanumeric key number exactly as it is printed on the label in the area located at the bottom of the dialog box.
6. Attach the label to the rear panel over the existing label.

## Verify the Option Upgrade

Complete the following steps to verify that the option upgrade was successful:

1. Restart the instrument.
2. Select the Help menu, and then select **About Tektronix AWG**.
3. Verify that the option that you installed is listed in the Options box. (See Figure 3 on page 4.)

The following example shows that Option 01 is installed on the instrument.




Figure 3: About Tektronix AWG

**End of Document**